Cold finger

Beamline: SXRMB

Contact: Mohsen Shakouri

Provides sample cooling down to 77 K (-196 ºC) using liquid nitrogen for low temperature x-ray absorption measurements.