

Helium cryostat #3

Beamline: HXMA

Contact: Ning Chen

Sample cooling down to ~ 7 K (-266°C) for low temperature x-ray absorption studies. ~ 5 K (-268°C) is possible with advanced notice.

