

# Helium cryostat #2

**Beamline:** BioXAS

**Contact:** Gosia Korbas

Provides sample cooling for temperature dependent x-ray absorption measurements on the BioXAS Side branch. Cooling down to (-193 °C) with liquid nitrogen is the default mode. Cooling down to 10 K (-263 °C) with liquid helium is also possible, given advanced notice to beamline staff.

