

Helium cryostat #1

Beamline: BioXAS

Contact: Gosia Korbias

Provides sample cooling for temperature dependent x-ray absorption measurements on the BioXAS Main branch. Cooling down to 80 K (-193 °C) with liquid nitrogen is the default mode. Cooling down to 10 K (-263 °C) with liquid helium is also possible, given advanced notice to beamline staff.

