

# X-PEEM conductive cooling stage

Beamline: SM

Contact: Jian Wang

Provides sample cooling down to 120 K (-153 °C) inside the X-PEEM endstation, enabling low temperature spectromicroscopy investigations.

