Humidity-stream

Beamline: CMCF

Contact: Michel Fodge

Provides sample relative humidity control (RH), from 0.30 – 0.995 at room temperature, by flowing a stream of humid air over the sample. The device can be installed on the CMCF BM or ID line with prior notice.

More information can be found in J. Sanchez-Weatherby et al., Acta Cryst. D65, 1237–1246 (2009), and on the CMCF website, https://cmcf.lightsource.ca/user-guide/humidity-control-xas/

